



BỘ Y TẾ
VIỆN PASTEUR TP. HCM

Institut Pasteur

KHOA XÉT NGHIỆM SINH HỌC LÂM SÀNG

252 Nam Kỳ Khởi Nghĩa, Quận 3 - Thành phố Hồ Chí Minh, Việt Nam
Điện thoại : (84.8) 38.297308 - 38.230352 - Fax : (84.8) 38.201882



VILAS Med 011

အိတ်စတိတ်

BACTERIAL EXAMINATION

(Testing antiseptic)

Code: 290917-4912

From : DAI NGUYEN LAM Co., Ltd
Address : 43 Street 4C, Ba Tanh, Commune 5, Binh Hung, Binh Chanh
Sample : STERIPLANT
Received on : The 29th of September 2017
Examined on : The 29th of September 2017

Material and Method:

Testing of sample concentration : 100%
Bacterial train : *Mycobacterium tuberculosis* ATCC 25177
Testing of bacterial concentration : 2.5×10^5 CFU/ml
Contact time : 5 minutes
Medium of culture : Lowenstein Jensen
Incubation : 30 days/37^oC


RESULT

Bacterial strain	After of contact sample 5 mins	
	Number of bacteria surviving (CFU/ml)	Disinfection percent (%)
<i>Mycobacterium tuberculosis</i> 2.5×10^5 CFU/ml	00	100

Conclusion: The result of testing STERIPLANT is as above.

Remark: This result is only valid on the sample.

Ho Chi Minh City, 29th of October, 2017



ThS. Vũ Lê Ngọc Lan